



Well-designed working table (luxury glass table for medical isntruments and medicine bottle)



■ ENT professional examination ■ Built-in soak sterilize endoscope barrel cup microscope



■ Constant Temperature Rinsing System



■ Liquid、Powder Spray gun



■ Contaminated instrument tank ■ Built-in waste tank





STANDARD EQUIPMENT:

| Main treatment table Size:770(L)*650(W)*805(H) (mm) | 1 |
|---|---|
| Subsidiary treatment table Size:1090(L)*500*(W)*805(H) (mm) | 1 |
| Liquid、Powder Spray gun(unbent3 & bent1) | 4 |
| Blowing gun | 1 |
| Suction gun | 1 |
| Laryngoscope pre-heater | 1 |
| LED illuminating Light | 1 |
| Compressor | 1 |
| Vacuum pump | 1 |
| Instrument tray | 2 |
| Medical bottle | 6 |
| Cotton cup | 4 |
| waste tray | 1 |
| Contaminated equipment placed tank | 1 |
| Built-in waste tank | 1 |
| Built-in endoscope barrel cup | 1 |
| Built-in LED cold light source | 1 |
| Film viewer | 1 |
| Blow-off equipment system(with warning system) | 1 |
| Doctor stool | 1 |
| Micro computer control system | 1 |

OPTION

Imaging system:

| 1CCD/3CCD camera |
|--|
| 15 inch LCD monitor |
| Computer (17inch LCD display/ professional endoscope software) |
| Printer |
| Fiber cable |
| built out hemia light/built out LED cold light source |
| Endoscope: |

Endoscope:

| 8mm, 70° /90° laryngoscope |
|-------------------------------------|
| microscope |
| Constant temperature rinsing system |
| |

SPECIFICATIONS:

| | voltage:AC230V 50Hz |
|---|---|
| 1 | power:1800W |
| i | Body size:181*70*80 (cm) |
| Ī | 01case weight:N.W/105kg G.W/151kg |
| Ī | 02case weight:N.W/116kg G.W/162kg |
| 1 | Positive pressure:1.2-2.0kgf/cm² Suction pressure:700mmHg |
| Ī | Main suction bottle:2500cc |
| 1 | Pre-heated device:450W |
| 1 | LED illuminating Light:9V 6W |